

# EMS System for Metropolitan Oklahoma City and Tulsa 2024 Medical Control Board Treatment Protocols



**EMS Section** 

Approved 11/08/23, Effective 1/15/24, replaces all prior versions

# **5E - TRANSCUTANEOUS PACING ADULT & PEDIATRIC**

### **PARAMEDIC**

### Indications:

- 1. Symptomatic 2nd Degree AV Block-Type II (Classic)
- 2. Symptomatic 3rd Degree AV Block (Complete)
- 3. Symptomatic Bradycardia in Acute Coronary Syndrome in preference to atropine use
- 4. Symptomatic Bradycardia unresponsive to non-electrical interventions
- 5. Symptomatic Bradycardia in pediatric patients (when approved by OLMCP consultation)

## Contraindications:


1. Asymptomatic Bradycardia


### Technique:

# (Physio-Control LifePak® 15):

- 1. Maintain standard ECG monitoring using electrodes/cable.
- 2. Apply Quik–Combo™pads in anterior/posterior chest wall location illustrated in (Figure 1). Excessive diaphoresis may require drying and/or excessive chest hair may require partial removal to achieve appropriate pad-chest wall adhesion.
- 3. Connect Quik-Combo™ pad set to LifePak® monitor/defibrillator via attached cable.
- 4. Advise patient of impending therapy. Administer sedation if patient condition allows, adults to receive 2-5 mg midazolam IVP as individual patient weight and hemodynamics dictate.
- 5. Power on the pacing function by pressing the "PACER" button (Figure 2).
- 6. Confirm ECG rhythm is sensed by Quik-Combo™ pads, looking for triangular "sense markers" marking QRS complexes (Figure 3). If sense markers do not appear, check for correct Quik-Combo™ pad attachment to LifePak monitor/defibrillator. If sense markers are inconsistently tracking QRS complexes and/or tracking T waves, adjust ECG size or select alternate monitoring lead to achieve correct QRS complex tracking.









# EMS System for Metropolitan Oklahoma City and Tulsa 2024 Medical Control Board Treatment Protocols






Approved 11/08/23, Effective 1/15/24, replaces all prior versions

# PROTOCOL 5E: Transcutaneous Pacing, Adult & Pediatric, cont.

## Technique (cont):

- 7. Set pacing **rate** at 60 paces per minute (adults) either by pressing the "RATE" switch up arrow to increase rate or down arrow to decrease rate or by rotating the "SPEED DIAL" knob (Figure 4). The "RATE" switch will allow changes in 10 paces per minute increments; the "SPEED DIAL" knob will allow changes in 5 paces per minute increments.
- 8. Set pacing **current** at minimum level achieving electrical AND mechanical capture. Deliver electrical pacing current either by pressing the "CURRENT" switch up arrow to increase milliAmp (mA) current or down arrow to decrease mA current or by rotating the "SPEED DIAL" knob (Figure 5). The "CURRENT" switch will allow changes in 10 mA increments; the "SPEED DIAL" knob will allow changes in 5 mA increments.
- 9. Pressing the "PAUSE" button will cause the set pacing rate to decrease by 25% (eg. rate of 60 paces per minute changes to rate of 45 paces per minute) while it is being depressed. This function should not be used without directive from OLMC.
- 10. If pacing therapy termination is required, power off the pacing function by pressing the "PACER" button (Figure 2).

# Figure 4



# Pacing-Related Considerations: (Physio-Control LifePak® 15):

- 1. In the event of ventricular fibrillation or pulseless ventricular tachycardia, pressing the yellow "CHARGE" button will automatically stop the pacing function. Proceed with defibrillation.
- 2. If the monitor displays "ECG LEAD OFF" during transcutaneous pacing, pacing automatically switches to non - demand and continues at the fixed rate until the ECG lead(s) is reattached. During non – demand pacing, the pacemaker delivers pulses at the set pace rate regardless of any intrinsic beats that the patient may have. The monitor continues to display the pacing rate and the current. To reestablish demand pacing, reattach the ECG lead(s).
- 3. If the Quik-Combo™ electrodes detach during pacing, the monitor will display "CONNECT **ELECTRODES"** and "PACING STOPPED" messages and sound an alarm. The set pacing rate is maintained, but the current resets to 0 mA. Reattaching the Quik-Combo™ electrodes silences the alarm and removes the messages. The current remains at 0 mA until manually adjusted as described above.



# EMS System for Metropolitan Oklahoma City and Tulsa **2024 Medical Control Board Treatment Protocols**



**EMS Section** 

Approved 11/08/23, Effective 1/15/24, replaces all prior versions

# PROTOCOL 5E: Transcutaneous Pacing, Adult & Pediatric, cont.

## Pacing-Related Considerations (cont):

- Proper electrical capture is displayed by depolarization of the ventricles, reflected as a wide QRS, followed by a distinct, broad T wave (Figures 6 & 7). Absence of these findings immediately following pacing spikes generally indicates failure of consistent electrical and mechanical capture (Figures 8 & 9).
- With transcutaneous pacing, it may be difficult to see the paced QRS complex due to washout from the pacing stimulus. It is imperative to confirm capture by a physiologic measure such as a pulse.

